

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 03155 Calib. Date: 12/04/2023 Calib. No.: 00049
Certification File No.: 03156 Cert. Date: 12/04/2023 Cert. No.: 00043
Linearity File No.: 03157 Lin. Date: 12/04/2023 Lin. No.: 00042
Solution File No.: 03172 Soln. Date: 01/01/2024 Soln. No.: 00382
Sequential File No.: 03172 File Date: 01/01/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 04/04/2024
Solution Control Lot: 22110 Bottle No.: 0934

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:15S		
Control 1 EC	0.103%	20:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	20:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:16S		
Control 2 EC	0.101%	20:17S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	20:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:18S		
Control 3 EC	0.101%	20:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	20:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ADAIR

First Name: CRAIG

MI: M

Signature: Sgt. C. R. #200

Badge No.: 200

Date: 01/01/2024