

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	03037	Calib. No.: 00041
Certification File No.:	03038	Cert. No.: 00037
Linearity File No.:	03039	Lin. No.: 00036
Solution File No.:	03042	Soln. No.: 00345
Sequential File No.:	03042	File Date: 12/31/2020

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUN S3-0340
Control Solution %: 0.100%		Expires: 09/02/2022
Solution Control Lot: 20390		Bottle No.: 0074

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:05S		
Control 1 EC	0.099%	20:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:06S		
Control 2 EC	0.098%	20:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:08S		
Control 3 EC	0.099%	20:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

M. J. Mireles # 219

Badge No.: 219

Date: 12/31/2020

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 03015 Calib. Date: 06/25/2020 Calib. No.: 00040
Certification File No.: 03016 Cert. Date: 06/25/2020 Cert. No.: 00036
Linearity File No.: 03017 Lin. Date: 06/25/2020 Lin. No.: 00035
Solution File No.: 03031 Soln. Date: 11/13/2020 Soln. No.: 00343
Sequential File No.: 03031 File Date: 11/13/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 1282

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:28S		
Control 1 EC	0.102%	20:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:30S		
Control 2 EC	0.101%	20:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	20:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:31S		
Control 3 EC	0.101%	20:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	20:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: *M. A. Mireles #219*

Badge No.: 219

Date: 11/13/2020

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 03015 Calib. Date: 06/25/2020 Calib. No.: 00040
Certification File No.: 03016 Cert. Date: 06/25/2020 Cert. No.: 00036
Linearity File No.: 03017 Lin. Date: 06/25/2020 Lin. No.: 00035
Solution File No.: 03027 Soln. Date: 10/10/2020 Soln. No.: 00342
Sequential File No.: 03027 File Date: 10/10/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 1368

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:26D		
Control 1 EC	0.101%	20:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	20:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:27D		
Control 2 EC	0.100%	20:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	20:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:29D		
Control 3 EC	0.100%	20:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: PASQUALE

First Name: JODY

MI:

Signature: _____



Badge No.: 199

Date: 10/10/2020

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
 Location: PENNSAUKEN TWSP. P.D.
 Calibration File No.: 03015 Calib. Date: 06/25/2020 Calib. No.: 00040
 Certification File No.: 03016 Cert. Date: 06/25/2020 Cert. No.: 00036
 Linearity File No.: 03017 Lin. Date: 06/25/2020 Lin. No.: 00035
 Solution File No.: 03024 Soln. Date: 08/29/2020 Soln. No.: 00341
 Sequential File No.: 03024 File Date: 08/29/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
 Control Solution %: 0.100% Expires: 09/09/2021
 Solution Control Lot: 19240 Bottle No.: 1326

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:07D		
Control 1 EC	0.100%	04:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:08D		
Control 2 EC	0.099%	04:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:09D		
Control 3 EC	0.099%	04:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	04:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: TIMOTHY

MI:

Signature: _____

Timothy Mireles # 219

Badge No.: 219

Date: 08/29/2020

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
 Location: PENNSAUKEN TWSP. P.D.
 Calibration File No.: 03015 Calib. Date: 06/25/2020 Calib. No.: 00040
 Certification File No.: 03016 Cert. Date: 06/25/2020 Cert. No.: 00036
 Linearity File No.: 03017 Lin. Date: 06/25/2020 Lin. No.: 00035
 Solution File No.: 03021 Soln. Date: 07/26/2020 Soln. No.: 00340
 Sequential File No.: 03021 File Date: 07/26/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
 Control Solution %: 0.100% Expires: 09/09/2021
 Solution Control Lot: 19240 Bottle No.: 1308

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:49D		
Control 1 EC	0.103%	03:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	03:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:50D		
Control 2 EC	0.102%	03:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:52D		
Control 3 EC	0.102%	03:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	03:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 191

Date: 07/26/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02976	Calib. Date: 01/16/2020
Certification File No.:	02977	Calib. No.: 00039
Linearity File No.:	02978	Cert. Date: 01/16/2020
Solution File No.:	03013	Cert. No.: 00035
Sequential File No.:	03013	Lin. Date: 01/16/2020
		Lin. No.: 00034
		Soln. Date: 06/11/2020
		Soln. No.: 00338
		File Date: 06/11/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1388

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:35D		
Control 1 EC	0.101%	04:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:36D		
Control 2 EC	0.099%	04:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	04:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:37D		
Control 3 EC	0.098%	04:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HURLEY

First Name: R

MI:

Signature: _____

P. R. H. # 205

Badge No.: 205

Date: 06/11/2020

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
 Location: PENNSAUKEN TWSP. P.D.
 Calibration File No.: 02976 Calib. Date: 01/16/2020 Calib. No.: 00039
 Certification File No.: 02977 Cert. Date: 01/16/2020 Cert. No.: 00035
 Linearity File No.: 02978 Lin. Date: 01/16/2020 Lin. No.: 00034
 Solution File No.: 03010 Soln. Date: 05/11/2020 Soln. No.: 00337
 Sequential File No.: 03010 File Date: 05/11/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
 Control Solution %: 0.100% Expires: 09/09/2021
 Solution Control Lot: 19240 Bottle No.: 1265

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:25D		
Control 1 EC	0.098%	16:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:27D		
Control 2 EC	0.096%	16:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:28D		
Control 3 EC	0.097%	16:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____

Peter Fabiani 187

Badge No.: 187

Date: 05/11/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02976	Calib. Date: 01/16/2020
Certification File No.:	02977	Calib. No.: 00039
Linearity File No.:	02978	Cert. Date: 01/16/2020
Solution File No.:	03009	Cert. No.: 00035
Sequential File No.:	03009	Lin. Date: 01/16/2020
		Lin. No.: 00034
		Soln. Date: 05/11/2020
		Soln. No.: 00336
		File Date: 05/11/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1271

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:15D		
Control 1 EC	-.---%	10:15D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	10:15D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____



Badge No.: 187

Date: 05/11/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02976	Calib. Date: 01/16/2020
Certification File No.:	02977	Calib. No.: 00039
Linearity File No.:	02978	Cert. Date: 01/16/2020
Solution File No.:	03003	Cert. No.: 00035
Sequential File No.:	03003	Lin. Date: 01/16/2020
		Lin. No.: 00034
		Soln. Date: 04/30/2020
		Soln. No.: 00335
		File Date: 04/30/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1344

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:41D		
Control 1 EC	0.099%	03:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:43D		
Control 2 EC	0.098%	03:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:44D		
Control 3 EC	0.098%	03:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HURLEY

First Name: R

MI:

Signature: _____

P.H. R.H.S. #205

Badge No.: 205

Date: 04/30/2020

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
 Location: PENNSAUKEN TWSP. P.D.
 Calibration File No.: 02976 Calib. Date: 01/16/2020 Calib. No.: 00039
 Certification File No.: 02977 Cert. Date: 01/16/2020 Cert. No.: 00035
 Linearity File No.: 02978 Lin. Date: 01/16/2020 Lin. No.: 00034
 Solution File No.: 03000 Soln. Date: 03/29/2020 Soln. No.: 00334
 Sequential File No.: 03000 File Date: 03/29/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
 Control Solution %: 0.100% Expires: 09/09/2021
 Solution Control Lot: 19240 Bottle No.: 1356

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:14S		
Control 1 EC	0.098%	15:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:15S		
Control 2 EC	0.097%	15:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17S		
Control 3 EC	0.097%	15:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	15:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HURLEY

First Name: R

MI:

Signature: _____

J+I. R. HZ #205

Badge No.: 205

Date: 03/29/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02976	Calib. Date: 01/16/2020
Certification File No.:	02977	Calib. No.: 00039
Linearity File No.:	02978	Cert. Date: 01/16/2020
Solution File No.:	02996	Cert. No.: 00035
Sequential File No.:	02996	Lin. Date: 01/16/2020
		Lin. No.: 00034
		Soln. Date: 03/22/2020
		Soln. No.: 00333
		File Date: 03/22/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0308

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:21S		
Control 1 EC	0.098%	06:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22S		
Control 2 EC	0.097%	06:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23S		
Control 3 EC	0.097%	06:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

Mireles #219

Badge No.: 219

Date: 03/22/2020

Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02976 Calib. Date: 01/16/2020 Calib. No.: 00039
Certification File No.: 02977 Cert. Date: 01/16/2020 Cert. No.: 00035
Linearity File No.: 02978 Lin. Date: 01/16/2020 Lin. No.: 00034
Solution File No.: 02991 Soln. Date: 03/02/2020 Soln. No.: 00332
Sequential File No.: 02991 File Date: 03/02/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 1361

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:49S		
Control 1 EC	0.100%	02:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:51S		
Control 2 EC	0.097%	02:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:52S		
Control 3 EC	0.097%	02:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: _____

PTL. BWA #227

Badge No.: 227

Date: 03/02/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02976	Calib. Date: 01/16/2020
Certification File No.:	02977	Calib. No.: 00039
Linearity File No.:	02978	Cert. Date: 01/16/2020
Solution File No.:	02984	Cert. No.: 00035
Sequential File No.:	02984	Lin. Date: 01/16/2020
		Lin. No.: 00034
		Soln. Date: 02/14/2020
		Soln. No.: 00331
		File Date: 02/14/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0307

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:58S		
Control 1 EC	0.098%	22:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:59S		
Control 2 EC	0.098%	23:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:01S		
Control 3 EC	0.097%	23:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____

L H Henkel 149

Badge No.: 149

Date: 02/14/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02905	Calib. Date: 08/26/2019
Certification File No.:	02906	Calib. No.: 00038
Linearity File No.:	02907	Cert. Date: 08/26/2019
Solution File No.:	02974	Cert. No.: 00034
Sequential File No.:	02974	Lin. Date: 08/26/2019
		Lin. No.: 00033
		Soln. Date: 01/11/2020
		Soln. No.: 00329
		File Date: 01/11/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1379

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:36S		
Control 1 EC	0.102%	20:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:37S		
Control 2 EC	0.101%	20:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:39S		
Control 3 EC	0.100%	20:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

PH. of Mireles #219

Badge No.: 219

Date: 01/11/2020