

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02875	Calib. Date: 02/09/2022
Certification File No.:	02876	Cert. No.: 00042
Linearity File No.:	02877	Cert. Date: 02/09/2022
Solution File No.:	02881	Lin. No.: 00038
Sequential File No.:	02881	Soln. Date: 03/05/2022
		Soln. No.: 00316
		File Date: 03/05/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0609

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:36S		
Control 1 EC	0.101%	23:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:37S		
Control 2 EC	0.099%	23:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:39S		
Control 3 EC	0.098%	23:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

M. Mireles 2/19

Badge No.: 219

Date: 03/05/2022