

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
PENNSAUKEN TWSP. P.D.

Serial No.: ARUM-0051

Location:
 Calibration File No.: 03114 Calib. Date: 12/19/2022
 Certification File No.: 03115 Cert. Date: 12/19/2022
 Linearity File No.: 03116 Lin. Date: 12/19/2022
 Solution File No.: 03122 Soln. Date: 01/18/2023
 Sequential File No.: 03122 File Date: 01/18/2023

Calib. No.: 00046
 Cert. No.: 00042
 Lin. No.: 00041
 Soln. No.: 00370

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 21270

Model No.: CU-34

Serial No.: DDUN S3-0339
 Expires: 07/21/2023
 Bottle No.: 0757

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:26S		
Control 1 EC	---	01:26S	34.0°C	CTRL GAS SUPPLY
Control 1 IR	---	01:26S	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	---	---		
Control 2 EC	---	---	N/A	
Control 2 IR	---	---	N/A	
Ambient Air Blank	---	---		
Control 3 EC	---	---	N/A	
Control 3 IR	---	---	N/A	
Ambient Air Blank	---	---		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISP) Chief Forensic Scientist.

Changed By:

Last Name: ADAIR

First Name: CRAIG

MI: M

Signature: Sgt. C. R. Farrow

Badge No.: 200

Date: 01/18/2023

Calibrating Unit New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	03114	Calib. Date:	12/19/2022
Certification File No.:	03115	Cert. Date:	12/19/2022
Linearity File No.:	03116	Lin. Date:	12/19/2022
Solution File No.:	03129	Soln. Date:	02/24/2023
Sequential File No.:	03129	File Date:	02/24/2023

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0341
Control Solution %:	0.100%			Expires:	04/04/2024
Solution Control Lot:	22110			Bottle No.:	1029

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:13S		
Control 1 EC	0.101%	04:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:15S		
Control 2 EC	0.099%	04:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:16S		
Control 3 EC	0.099%	04:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:		First Name:	CRAIG	MI:	M
Last Name:	ADAIR			Badge No.:	200
Signature:				Date:	02/24/2023

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
PENNSAUKEN TWSP. P.D.

Serial No.: ARUM-0051

Location: 03114
 Calibration File No.: 03114
 Certification File No.: 03115
 Linearity File No.: 03116
 Solution File No.: 03134
 Sequential File No.: 03134

Calib. Date: 12/19/2022
 Cert. No.: 00046
 Cert. No.: 00042
 Lin. Date: 12/19/2022
 Lin. No.: 00041
 Soln. Date: 03/27/2023
 Soln. No.: 00374
 File Date: 03/27/2023

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 22110

Model No.: CU-34
 Serial No.: DDUN S3-0341
 Expires: 04/04/2024
 Bottle No.: 0941

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:52S		
Control 1 EC	0.100%	19:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54S		
Control 2 EC	0.097%	19:54S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:55S		
Control 3 EC	0.098%	19:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:56S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARTINEZ

First Name: JUAN

MI: C

Signature: 

Badge No.: 27223

Date: 03/27/2023

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
PENNSAUKEN TWSP. P.D.

Serial No.: ARUM-0051

Location: 03114
Calibration File No.: 03114
Certification File No.: 03115
Linearity File No.: 03116
Solution File No.: 03137
Sequential File No.: 03137

Calib. Date: 12/19/2022
Cert. Date: 12/19/2022
Lin. Date: 12/19/2022
Soln. Date: 04/22/2023
File Date: 04/22/2023

Calib. No.: 00046
Cert. No.: 00042
Lin. No.: 00041
Soln. No.: 00375

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 22110

Model No.: CU-34

Serial No.: DDUN S3-0341
Expires: 04/04/2024
Bottle No.: 0933

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:38D		
Control 1 EC	0.099%	08:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:40D		
Control 2 EC	0.097%	08:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41D		
Control 3 EC	0.098%	08:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:42D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISP) Chief Forensic Scientist.

Changed By:

Last Name: MARTINEZ

First Name: JUAN

MI: C

Signature: 

Badge No.: 27223

Date: 04/22/2023

Calibrating Unit New Standard Solution Report

Equipment

Alcolest 7110 MKIII-C
PENNSAUKEN TWSP. P.D.

Serial No.: ARUM-0051

Location: 03114
Calibration File No.: 03114
Certification File No.: 03115
Linearity File No.: 03116
Solution File No.: 03148
Sequential File No.: 03148

Calib. Date: 12/19/2022
Cert. Date: 12/19/2022
Lin. Date: 12/19/2022
Soln. Date: 05/28/2023
File Date: 05/28/2023

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 22110

Model No.: CU-34

Serial No.: DDUN S3-0341
Expires: 04/04/2024
Bottle No.: 0841

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:53D		
Control 1 EC	0.098%	19:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54D		
Control 2 EC	0.097%	19:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:56D		
Control 3 EC	0.097%	19:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:57D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcolest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PEREZ

First Name: EDDIE

MI: --

Signature: 

#231

Badge No.: 27237
Date: 05/28/2023

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
PENNSAUKEN TWS/P. P.D.

Serial No.: ARUM-0051

Location: PENNSAUKEN TWS/P. P.D.
 Calibration File No.: 03078 Calib. Date: 02/09/2022
 Certification File No.: 03079 Cert. Date: 02/09/2022
 Linearity File No.: 03080 Lin. Date: 02/09/2022
 Solution File No.: 03086 Soln. Date: 05/14/2022
 Sequential File No.: 03086 File Date: 05/14/2022

Calib. No.: 00044
 Cert. No.: 00040
 Lin. No.: 00039
 Soln. No.: 00362

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 21270

Model No.: CU-34

Serial No.: DDUN S3-0341
 Expires: 07/21/2023
 Bottle No.: 0155

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:43D		
Control 1 EC	0.100%	21:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:45D		
Control 2 EC	0.098%	21:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:46D		
Control 3 EC	0.098%	21:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ADAIR

First Name: CRAIG

MI: M

Signature:

Craig ADAIR #200

Badge No.: 200
 Date: 05/14/2022

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
 Location: PENNSAUKEN TWSP. P.D.

Calibration File No.:	03078	Calib. Date:	02/09/2022	Calib. No.:	00044
Certification File No.:	03079	Cert. Date:	02/09/2022	Cert. No.:	00040
Linearity File No.:	03080	Lin. Date:	02/09/2022	Lin. No.:	00039
Solution File No.:	03089	Soln. Date:	06/11/2022	Soln. No.:	00363
Sequential File No.:	03089	File Date:	06/11/2022		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0341
Control Solution %:	0.100%	Expires:	07/21/2023	Bottle No.:	0097
Solution Control Lot:	21270				

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:06D		
Control 1 EC	0.098%	20:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	20:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:08D		
Control 2 EC	0.097%	20:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	20:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09D		
Control 3 EC	0.097%	20:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	20:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NISF) Chief Forensic Scientist.

Changed By:	
Last Name: ADAIR	First Name: CRAIG
Signature: <u><i>Sgt-C. BAIRD</i></u>	Badge No.: 200
	Date: 06/11/2022
	MI: M

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP, P.D.		
Calibration File No.:	03090	Calib. Date:	07/11/2022
Certification File No.:	03091	Cert. Date:	07/11/2022
Linearity File No.:	03080	Lin. Date:	02/09/2022
Solution File No.:	03089	Soln. Date:	06/11/2022
Sequential File No.:	03091	File Date:	07/11/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0340
Solution Control Lot:	21210	Expires:	06/16/2023
		Bottle No.:	1179

Function	Result	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:43D		
Control 1 EC	0.100%	09:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:44D		
Control 2 EC	0.099%	09:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45D		
Control 3 EC	0.099%	09:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:46D		

All tests within acceptable tolerance.

Coordinator

Last Name: WATSON

First Name: MATTHEW

MI: R

Badge No.: 7078

Date: 07/11/2022

Signature:



Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are willfully false, I am subject to punishment.

Alcotest 7110 Calibration Record

Serial No.: ARUM-0051

Equipment

Alcotest 7110 MKIII-C
PENNSAUKEN TWS.P. P.D.

Location: PENNSAUKEN TWS.P. P.D.
Calibration File No.: 03090
Certification File No.: 03079
Linearity File No.: 03080
Solution File No.: 03089
Sequential File No.: 03090

Calib. Date: 07/11/2022
Cert. Date: 02/09/2022
Lin. Date: 02/09/2022
Soln. Date: 06/11/2022
File Date: 07/11/2022

Calib. No.: 00045
Cert. No.: 00040
Lin. No.: 00039
Soln. No.: 00363

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 21210

Model No.: CU-34

Serial No.: DDUN S3-0340
Expires: 06/16/2023
Bottle No.: 1179

Coordinator

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature: Sgt. [Signature]

Badge No.: 7078
Date: 07/11/2022

*Black Key Temperature Probe Serial.....# DDMBP1-0119 MRM
*Digital NIST Temperature Measuring System Serial.....# 210216813 MRM

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are willfully false, I am subject to punishment.

Alcotest 7110 MKIII-C Calibration
NIST-Traceable Digital Thermometer Readings

Coordinator:

Sgt. Matthew R. Watson
Name

7078
Badge No.

Location:

Pennsauken Twp. P.D.
Agency

ARUM-0251
Alcotest Serial No.

Equipment:

210216813
Digital NIST Temperature Measuring System Serial No.

Simulator Solution Concentration	CU-34 Simulator Serial No.	Time Simulators Started to Heat	Time Temp. Reading Obtained	Temp. Reading on NIST Traceable Thermometer
0.04%	DDXD S3-0187	08:21p	09:32p	33.9°C
0.08%	DDRK S3-0015	08:21p	09:34p	34.0°C
0.10%	DDUN S3-0340	08:21p	09:35p	34.0°C
0.16%	DDRK S3-0006	08:21p	09:37p	33.9°C

Pursuant to law and the "Chemical Breath Testing Regulations" established at N.J.A.C. 13-51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity and consistent with the "Calibration Check Procedure for Alcotest 7110" as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on Alcotest 7110 MKIII-C instruments. Pursuant to and consistent with the current "Calibration Check Procedure for Alcotest 7110", I performed a Calibration Check Procedure on the Alcotest 7110 MKIII-C instrument identified on this certificate. Pursuant to the current "Calibration Check Procedure for Alcotest 7110", I used the Digital NIST-traceable Temperature Measuring System identified on this certificate to confirm that the temperatures of the 0.10%, 0.04%, 0.08%, and 0.16% Simulator Solutions used in the respective CU-34 Simulators identified on this certificate, were 34.0 degrees Celsius ± 0.2 degrees Celsius. I hereby certify that I truthfully recorded on this certificate the temperatures of each of the simulator solutions as shown on the Digital NIST-traceable Temperature Measuring System thermometer. I am aware that if any of the foregoing statements made by me are willfully false, I am subject to punishment.

Sgt.  #7078
Coordinator's Signature

0711112022
Date

Calibrating Unit New Standard Solution Report

Equipment

Alcocest 7110 MKIII-C
PENNSAUKEN TWSP. P.D.

Serial No.: ARUM-0051

Location:
Calibration File No.: 03090
Certification File No.: 03091
Linearity File No.: 03092
Solution File No.: 03093
Sequential File No.: 03093

Calib. Date: 07/11/2022
Cert. Date: 07/11/2022
Lin. Date: 07/11/2022
Soln. Date: 07/11/2022
File Date: 07/11/2022

Calib. No.: 00045
Cert. No.: 00041
Lin. No.: 00040
Soln. No.: 00364

Calibrating Unit: WEST
Control Solution %: 0.100%
Solution Control Lot: 21270

Model No.: CU-34

Serial No.: DDUN \$3-0340
Expires: 07/21/2023
Bottle No.: 0156

Function	Result	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:21D	34.0°C	*** TEST PASSED ***
Control 1 EC	0.100%	11:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22D	34.0°C	*** TEST PASSED ***
Control 2 EC	0.099%	11:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D	34.0°C	*** TEST PASSED ***
Control 3 EC	0.099%	11:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25D	34.0°C	*** TEST PASSED ***

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcocest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number:

DDU5P2-142

MEM

Changed By:

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature:

Sgt. Matthew Watson

Badge No.: 7078
Date: 07/11/2022

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PENNSAUKEN TWSP. P.D.

Serial No.: ARUM-0051

Calibration File No.: 03090 Calib. Date: 07/11/2022 Calib. No.: 00045
Certification File No.: 03091 Cert. Date: 07/11/2022 Cert. No.: 00041
Linearity File No.: 03092 Lin. Date: 07/11/2022 Lin. No.: 00040
Solution File No.: 03096 Soln. Date: 09/08/2022 Soln. No.: 00365
Sequential File No.: 03096 File Date: 09/08/2022

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 21270

Model No.: CU-34

Serial No.: DDUN S3-0340
Expires: 07/21/2023
Bottle No.: 0547

Function	Result	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:27D		
Control 1 EC	0.099%	23:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:28D		
Control 2 EC	0.097%	23:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:30D		
Control 3 EC	0.099%	23:30D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:31D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ADAJR

First Name: CRAIG

MI: M

Signature: 

Badge No.: 200

Date: 09/08/2022

Calibrating Unit New Standard Solution Report

Equipment

Alcoltest 7110 MKIII-C
PENNSAUKEN TWSP, P.D.

Serial No.: ARUM-0051

Location:
Calibration File No.: 03090
Certification File No.: 03091
Linearity File No.: 03092
Solution File No.: 03100
Sequential File No.: 03100

Calib. Date: 07/11/2022
Cert. Date: 07/11/2022
Lin. Date: 07/11/2022
Soln. Date: 10/10/2022
File Date: 10/10/2022

Calib. No.: 00045
Cert. No.: 00041
Lin. No.: 00040
Soln. No.: 00366

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 21270

Model No.: CU-34

Serial No.: DDUN S3-0340
Expires: 07/21/2023
Bottle No.: 0157

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:52D		
Control 1 EC	0.099%	20:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:53D		
Control 2 EC	0.097%	20:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:54D		
Control 3 EC	0.098%	20:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcoltest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ADAIR

First Name: CRAIG

MI: M

Signature:

Sgt. C. AD #200

Badge No.: 200
Date: 10/10/2022

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 12/19/94 Notices)
Dräger, Inc.

Model: ALCOTEST CU34

Model: MARK IIA

X-Cal 2000 (Alcosim)

Other: _____

Serial Number:

DDXDS3-01871

Certification Date:

4.5.22

Technician:

MB

Re-Certification Due Date:

4.5.23

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 12/19/94 Notices)
Dräger, Inc.

Model: ALCOTEST CU34

Model: MARK IIA

X-Cal 2000 (Alcosim)

Other: _____

Serial Number:

DDRKS3-DDIS

Certification Date:

4.5.22

Technician:

MB

Re-Certification Due Date:

4.5.23

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R.: Vol. 59 No. 249 12/19/94 Notices)

Dräger, Inc.

Serial Number:

- Model: ALCOTEST CU34
- Model: MARK IIA
- X-Cal 2000 (Alcosim)
- Other: _____

DDRHS3-0006

Certification Date:

4.1.22

Technician:

MB

Re-Certification Due Date:

4.1.23

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe: DDMBP1-0119 Certification Date: 4.1.22 Next Certification Due: 4.1.23

Probe Value:

103

Dräger, Inc.

MB



Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:

Certification Date:

Next Certification Due:

DDUJPA-142

06-15-22

06-15-23

Probe Value:

105

Dräger, Inc.

MR / DH



Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Dräger, Inc.

Model: ALCOTEST CU34

Model: MARK IIIA

X-Cal 2000 (Alcosim)

Other: _____

Serial Number:

DDUN S3-0340

Certification Date:

06-15-22

Technician:

MB / DH

Re-Certification Due Date:

06-15-23

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:

DDUJPA-142

Certification Date:

06-15-22

Next Certification Due:

06-15-23

Probe Value:

105

Dräger, Inc.

MB / DH