

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02175	Calib. Date:	01/11/2017
Certification File No.:	02176	Cert. Date:	01/11/2017
Linearity File No.:	02177	Lin. Date:	01/11/2017
Solution File No.:	02188	Soln. Date:	02/13/2017
Sequential File No.:	02188	File Date:	02/13/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0341
Solution Control Lot:	16330	Expires:	10/28/2018
		Bottle No.:	0022

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:35S		
Control 1 EC	0.101%	08:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:36S		
Control 2 EC	0.100%	08:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:38S		
Control 3 EC	0.101%	08:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MOLES

First Name: VITO

MI: M

Signature: \_\_\_\_\_

Badge No.: 171

Date: 02/13/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02175	Calib. Date: 01/11/2017
Certification File No.:	02176	Calib. No.: 00030
Linearity File No.:	02177	Cert. Date: 01/11/2017
Solution File No.:	02199	Cert. No.: 00026
Sequential File No.:	02199	Lin. Date: 01/11/2017
		Lin. No.: 00026
		Soln. Date: 03/11/2017
		Soln. No.: 00229
		File Date: 03/11/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0137

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:19S		
Control 1 EC	0.100%	08:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:20S		
Control 2 EC	0.099%	08:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:22S		
Control 3 EC	0.100%	08:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SAMPSON

First Name: SHAUN

MI: .

Signature: \_\_\_\_\_



Badge No.: 142

Date: 03/11/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02175	Calib. Date: 01/11/2017
Certification File No.:	02176	Calib. No.: 00030
Linearity File No.:	02177	Cert. Date: 01/11/2017
Solution File No.:	02217	Cert. No.: 00026
Sequential File No.:	02217	Lin. Date: 01/11/2017
		Lin. No.: 00026
		Soln. Date: 04/08/2017
		Soln. No.: 00230
		File Date: 04/08/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0537

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:37D		
Control 1 EC	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 2 EC	0.100%	09:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		
Control 3 EC	0.099%	09:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SAMPSON

First Name: OFFICER

MI: .

Signature: \_\_\_\_\_

*J. Sampson #142*

Badge No.: 142

Date: 04/08/2017

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02175 Calib. Date: 01/11/2017 Calib. No.: 00030  
Certification File No.: 02176 Cert. Date: 01/11/2017 Cert. No.: 00026  
Linearity File No.: 02177 Lin. Date: 01/11/2017 Lin. No.: 00026  
Solution File No.: 02229 Soln. Date: 05/07/2017 Soln. No.: 00231  
Sequential File No.: 02229 File Date: 05/07/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341  
Control Solution %: 0.100% Expires: 10/28/2018  
Solution Control Lot: 16330 Bottle No.: 0134

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:24D		
Control 1 EC	0.100%	00:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:25D		
Control 2 EC	0.099%	00:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:27D		
Control 3 EC	0.099%	00:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_



Badge No.: 187

Date: 05/07/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02175	Calib. Date:	01/11/2017
Certification File No.:	02176	Cert. Date:	01/11/2017
Linearity File No.:	02177	Lin. Date:	01/11/2017
Solution File No.:	02241	Soln. Date:	05/27/2017
Sequential File No.:	02241	File Date:	05/27/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0341
Solution Control Lot:	16330	Expires:	10/28/2018
		Bottle No.:	0023

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:54D		
Control 1 EC	0.098%	05:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:55D		
Control 2 EC	0.098%	05:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	05:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:56D		
Control 3 EC	0.098%	05:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_

Badge No.: 187

Date: 05/27/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02175	Calib. Date: 01/11/2017
Certification File No.:	02176	Calib. No.: 00030
Linearity File No.:	02177	Cert. Date: 01/11/2017
Solution File No.:	02255	Cert. No.: 00026
Sequential File No.:	02255	Lin. Date: 01/11/2017
		Lin. No.: 00026
		Soln. Date: 06/19/2017
		Soln. No.: 00233
		File Date: 06/19/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0246

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:29D		
Control 1 EC	0.100%	21:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:30D		
Control 2 EC	0.100%	21:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:31D		
Control 3 EC	0.100%	21:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_

*Fabiani P 187*

Badge No.: 187

Date: 06/19/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02175	Calib. Date: 01/11/2017
Certification File No.:	02176	Calib. No.: 00030
Linearity File No.:	02177	Cert. Date: 01/11/2017
Solution File No.:	02265	Cert. No.: 00026
Sequential File No.:	02265	Lin. Date: 01/11/2017
		Lin. No.: 00026
		Soln. Date: 07/02/2017
		Soln. No.: 00234
		File Date: 07/02/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0322

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:28D		
Control 1 EC	0.099%	23:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:29D		
Control 2 EC	0.098%	23:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:30D		
Control 3 EC	0.099%	23:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SOTO

First Name: JESUS

MI: -

Signature: \_\_\_\_\_



Badge No.: 198

Date: 07/02/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02271	Calib. Date:	07/11/2017
Certification File No.:	02272	Cert. Date:	07/11/2017
Linearity File No.:	02273	Lin. Date:	07/11/2017
Solution File No.:	02283	Soln. Date:	08/04/2017
Sequential File No.:	02283	File Date:	08/04/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0340
Control Solution %:	0.100%			Expires:	10/28/2018
Solution Control Lot:	16330			Bottle No.:	0347

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:13D		
Control 1 EC	0.099%	08:14D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15D		
Control 2 EC	0.098%	08:15D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16D		
Control 3 EC	0.098%	08:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KERN

First Name: RAYMOND

MI: T

Signature: \_\_\_\_\_



Badge No.: 181

Date: 08/04/2017



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02271	Calib. Date: 07/11/2017
Certification File No.:	02272	Calib. No.: 00031
Linearity File No.:	02273	Cert. Date: 07/11/2017
Solution File No.:	02297	Cert. No.: 00027
Sequential File No.:	02297	Lin. Date: 07/11/2017
		Lin. No.: 00027
		Soln. Date: 08/20/2017
		Soln. No.: 00237
		File Date: 08/20/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0340
Control Solution %:	0.100%		Expires: 03/20/2019
Solution Control Lot:	17110		Bottle No.: 0960

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:07D		
Control 1 EC	0.101%	06:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:09D		
Control 2 EC	0.100%	06:09D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10D		
Control 3 EC	0.100%	06:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_



Badge No.: 187

Date: 08/20/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02271	Calib. Date: 07/11/2017
Certification File No.:	02272	Calib. No.: 00031
Linearity File No.:	02273	Cert. Date: 07/11/2017
Solution File No.:	02304	Cert. No.: 00027
Sequential File No.:	02304	Lin. Date: 07/11/2017
		Lin. No.: 00027
		Soln. Date: 09/16/2017
		Soln. No.: 00238
		File Date: 09/16/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0895

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:07D		
Control 1 EC	-.---%	21:07D	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	21:07D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_



Badge No.: 187

Date: 09/16/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02271	Calib. Date: 07/11/2017
Certification File No.:	02272	Calib. No.: 00031
Linearity File No.:	02273	Cert. Date: 07/11/2017
Solution File No.:	02305	Cert. No.: 00027
Sequential File No.:	02305	Lin. Date: 07/11/2017
		Lin. No.: 00027
		Soln. Date: 09/16/2017
		Soln. No.: 00239
		File Date: 09/16/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0906

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:21D		
Control 1 EC	0.100%	22:22D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:23D		
Control 2 EC	0.100%	22:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:24D		
Control 3 EC	0.100%	22:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_

Badge No.: 187

Date: 09/16/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02271	Calib. Date: 07/11/2017
Certification File No.:	02272	Calib. No.: 00031
Linearity File No.:	02273	Cert. Date: 07/11/2017
Solution File No.:	02314	Cert. No.: 00027
Sequential File No.:	02314	Lin. Date: 07/11/2017
		Lin. No.: 00027
		Soln. Date: 10/13/2017
		Soln. No.: 00240
		File Date: 10/13/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0832

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:06D		
Control 1 EC	0.100%	23:07D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:07D		
Control 2 EC	0.099%	23:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:09D		
Control 3 EC	0.100%	23:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_



Badge No.: 187

Date: 10/13/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02271	Calib. Date: 07/11/2017
Certification File No.:	02272	Calib. No.: 00031
Linearity File No.:	02273	Cert. Date: 07/11/2017
Solution File No.:	02333	Cert. No.: 00027
Sequential File No.:	02333	Lin. Date: 07/11/2017
		Lin. No.: 00027
		Soln. Date: 11/15/2017
		Soln. No.: 00241
		File Date: 11/15/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0828

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:41S		
Control 1 EC	0.101%	04:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	04:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:43S		
Control 2 EC	0.100%	04:43S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	04:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:44S		
Control 3 EC	0.101%	04:45S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	04:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:45S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: 

Badge No.: 216

Date: 11/15/2017