

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02139	Calib. Date:	12/09/2013
Certification File No.:	02140	Cert. Date:	12/09/2013
Linearity File No.:	02141	Lin. Date:	12/09/2013
Solution File No.:	02153	Soln. Date:	01/19/2014
Sequential File No.:	02153	File Date:	01/19/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0341
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	0381

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:43S		
Control 1 EC	0.100%	20:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:44S		
Control 2 EC	0.098%	20:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:46S		
Control 3 EC	0.099%	20:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Pr J K #190

Badge No.: 190

Date: 01/19/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02139	Calib. Date: 12/09/2013
Certification File No.:	02140	Calib. No.: 00023
Linearity File No.:	02141	Cert. Date: 12/09/2013
Solution File No.:	02157	Cert. No.: 00020
Sequential File No.:	02157	Lin. Date: 12/09/2013
		Lin. No.: 00019
		Soln. Date: 02/11/2014
		Soln. No.: 00204
		File Date: 02/11/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0386

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:22S		
Control 1 EC	0.101%	23:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:23S		
Control 2 EC	0.099%	23:24S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:25S		
Control 3 EC	0.099%	23:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Pr Jh #190

Badge No.: 190

Date: 02/11/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02139	Calib. Date: 12/09/2013
Certification File No.:	02140	Calib. No.: 00023
Linearity File No.:	02141	Cert. Date: 12/09/2013
Solution File No.:	02165	Cert. No.: 00020
Sequential File No.:	02165	Lin. Date: 12/09/2013
		Lin. No.: 00019
		Soln. Date: 03/11/2014
		Soln. No.: 00205
		File Date: 03/11/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0387

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:16S		
Control 1 EC	0.101%	22:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:18S		
Control 2 EC	0.097%	22:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:20S		
Control 3 EC	0.098%	22:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

[Handwritten Signature] #190

Badge No.: 190

Date: 03/11/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02139	Calib. Date: 12/09/2013
Certification File No.:	02140	Calib. No.: 00023
Linearity File No.:	02141	Cert. Date: 12/09/2013
Solution File No.:	02171	Cert. No.: 00020
Sequential File No.:	02171	Lin. Date: 12/09/2013
		Lin. No.: 00019
		Soln. Date: 03/24/2014
		Soln. No.: 00206
		File Date: 03/24/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0389

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:18S		
Control 1 EC	0.099%	03:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:20S		
Control 2 EC	0.097%	03:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:21S		
Control 3 EC	0.098%	03:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

FBI P #187

Badge No.: 187

Date: 03/24/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02139	Calib. Date:	12/09/2013
Certification File No.:	02140	Cert. Date:	12/09/2013
Linearity File No.:	02141	Lin. Date:	12/09/2013
Solution File No.:	02175	Soln. Date:	04/19/2014
Sequential File No.:	02175	File Date:	04/19/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0341
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	0392

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:54D		
Control 1 EC	0.101%	04:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	04:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:55D		
Control 2 EC	0.099%	04:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:56D		
Control 3 EC	0.099%	04:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Joe Kuchmek #180

Badge No.: 190

Date: 04/19/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02139	Calib. Date: 12/09/2013
Certification File No.:	02140	Calib. No.: 00023
Linearity File No.:	02141	Cert. Date: 12/09/2013
Solution File No.:	02191	Cert. No.: 00020
Sequential File No.:	02191	Lin. Date: 12/09/2013
		Lin. No.: 00019
		Soln. Date: 05/24/2014
		Soln. No.: 00209
		File Date: 05/24/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0378

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:05D		
Control 1 EC	0.098%	05:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	05:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:06D		
Control 2 EC	0.097%	05:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	05:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:08D		
Control 3 EC	0.097%	05:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Badge No.: 190

Date: 05/24/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02139	Calib. Date:	12/09/2013
Certification File No.:	02140	Cert. Date:	12/09/2013
Linearity File No.:	02141	Lin. Date:	12/09/2013
Solution File No.:	02199	Soln. Date:	06/07/2014
Sequential File No.:	02199	File Date:	06/07/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0341
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	0398

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:43D		
Control 1 EC	0.097%	06:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	06:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:45D		
Control 2 EC	0.097%	06:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	06:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:46D		
Control 3 EC	0.097%	06:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:47D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: 

Badge No.: 190

Date: 06/07/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02200	Calib. Date:	06/09/2014
Certification File No.:	02201	Cert. Date:	06/09/2014
Linearity File No.:	02202	Lin. Date:	06/09/2014
Solution File No.:	02213	Soln. Date:	07/07/2014
Sequential File No.:	02213	File Date:	07/07/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0339
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	0400

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:26D		
Control 1 EC	0.100%	22:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	22:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:27D		
Control 2 EC	0.100%	22:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:29D		
Control 3 EC	0.100%	22:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:30D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature:  187

Badge No.: 187

Date: 07/07/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02200	Calib. Date:	06/09/2014
Certification File No.:	02201	Cert. Date:	06/09/2014
Linearity File No.:	02202	Lin. Date:	06/09/2014
Solution File No.:	02225	Soln. Date:	08/08/2014
Sequential File No.:	02225	File Date:	08/08/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0339
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	0394

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:51D		
Control 1 EC	-.---%	22:52D	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	22:52D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	---:---		
Control 2 EC	-.---%	---:---	N/A	
Control 2 IR	-.---%	---:---	N/A	
Ambient Air Blank	-.---%	---:---		
Control 3 EC	-.---%	---:---	N/A	
Control 3 IR	-.---%	---:---	N/A	
Ambient Air Blank	-.---%	---:---		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

Badge No.: 187

Date: 08/08/2014

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02200 Calib. Date: 06/09/2014 Calib. No.: 00024
Certification File No.: 02201 Cert. Date: 06/09/2014 Cert. No.: 00021
Linearity File No.: 02202 Lin. Date: 06/09/2014 Lin. No.: 00020
Solution File No.: 02226 Soln. Date: 08/09/2014 Soln. No.: 00214
Sequential File No.: 02226 File Date: 08/09/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 09/10/2015
Solution Control Lot: 13I119 Bottle No.: 0399

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:31D		
Control 1 EC	0.100%	01:31D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:32D		
Control 2 EC	0.099%	01:33D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:33D		
Control 3 EC	0.099%	01:34D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

 *187

Badge No.: 187

Date: 08/09/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02200	Calib. Date: 06/09/2014
Certification File No.:	02201	Calib. No.: 00024
Linearity File No.:	02202	Cert. Date: 06/09/2014
Solution File No.:	02238	Cert. No.: 00021
Sequential File No.:	02238	Lin. Date: 06/09/2014
		Lin. No.: 00020
		Soln. Date: 09/01/2014
		Soln. No.: 00215
		File Date: 09/01/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0395

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:19D		
Control 1 EC	0.099%	17:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20D		
Control 2 EC	0.099%	17:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:21D		
Control 3 EC	0.099%	17:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

[Handwritten Signature]

Badge No.: 187

Date: 09/01/2014

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02200 Calib. Date: 06/09/2014 Calib. No.: 00024
Certification File No.: 02201 Cert. Date: 06/09/2014 Cert. No.: 00021
Linearity File No.: 02202 Lin. Date: 06/09/2014 Lin. No.: 00020
Solution File No.: 02248 Soln. Date: 09/29/2014 Soln. No.: 00216
Sequential File No.: 02248 File Date: 09/29/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 06/20/2016
Solution Control Lot: 14F130 Bottle No.: 1311

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:51D		
Control 1 EC	0.102%	20:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	20:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:52D		
Control 2 EC	0.101%	20:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	20:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:54D		
Control 3 EC	0.101%	20:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	20:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

 187

Badge No.: 187

Date: 09/29/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02200	Calib. Date:	06/09/2014
Certification File No.:	02201	Cert. Date:	06/09/2014
Linearity File No.:	02202	Lin. Date:	06/09/2014
Solution File No.:	02256	Soln. Date:	10/27/2014
Sequential File No.:	02256	File Date:	10/27/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0339
Solution Control Lot:	14F130	Expires:	06/20/2016
		Bottle No.:	1301

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:35S		
Control 1 EC	0.101%	21:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	21:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:36S		
Control 2 EC	0.100%	21:37S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:37S		
Control 3 EC	0.100%	21:38S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature:  187

Badge No.: 187

Date: 10/27/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02200	Calib. Date:	06/09/2014
Certification File No.:	02201	Cert. Date:	06/09/2014
Linearity File No.:	02202	Lin. Date:	06/09/2014
Solution File No.:	02269	Soln. Date:	11/25/2014
Sequential File No.:	02269	File Date:	11/25/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0339
Solution Control Lot:	14F130	Expires:	06/20/2016
		Bottle No.:	1302

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:04S		
Control 1 EC	0.101%	19:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:05S		
Control 2 EC	0.100%	19:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:07S		
Control 3 EC	0.100%	19:07S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	19:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

FABIANI P 187

Badge No.: 187

Date: 11/25/2014